

CALIBRATION REPORT

Product	HISAC ULTIMA	Customer	Oasis Techno Engg Pvt. Ltd. Noida
Sr. No.	3015.00Y801	Calibrated on	25.01.2022
Report No.	CAL/SSP/AI191	Calibration Due on	24.01.2023

Reference Standard used for Calibration:

Calibrator	SCALE TCV	Calibrated by	ARAI Pune
Report Reference No.	ARAI/CAL/2109/1622	Calibration Due	22.09.2022

1.0 OBSERVATIONS:

Sr. No.	Parameter	Expected value		Measured Value		Remark
		Min	Max			
1.	Time (mSec)	99.850	100.150	R1 - 100.1 mSc	R2 - 100.1 mSc	OK/FAIL
				R3 - 100.1 mSc	R4 - 100.1 mSc	OK/FAIL
				Y1 - 100.1 mSc	Y2 - 100.1 mSc	OK/FAIL
				Y3 - 100.1 mSc	Y4 - 100.1 mSc	OK/FAIL
				B1 - 100.1 mSc	B2 - 100.1 mSc	OK/FAIL
				B3 - 100.1 mSc	B4 - 100.1 mSc	OK/FAIL
2.	Current (Amp)	0.995	1.007	CHANNEL C1 - 1.003 Amp		OK/FAIL
				CHANNEL C2 - 1.000 Amp		OK/FAIL
				CHANNEL C3 - 1.001 Amp		OK/FAIL
3.	Velocity (M/Sec)	0.94	1.06	CHANNEL TR - 1.01 M/Sec		OK/FAIL
				CHANNEL TY - 1.01 M/Sec		OK/FAIL
				CHANNEL TB - 0.99 M/Sec		OK/FAIL

2.0 BRIEF RESULTS:

The product under test complies with calibration specification indicated above and is verified to be within calibration.

Tested By	Certified By
	

