

CALIBRATION REPORT

Product	DCRM ULTIMA	Customer	OASIS Noida
Sr. No.	3021.00 W352	Calibrated on	30.12.2022
Report No.	CAL/ASS/AJ022	Calibration Due on	29.12.2023

Reference Standard used for Calibration:			
Calibrator	SHUNT SET 07	Calibrated by	Industrial Electronics & Instruments, Pune
Report Reference No	ILI/ST&M/CERT/11/22-23/ 0401	Calibration Due	26.07.2023

1.0 OBSERVATIONS:

Sr. No.	Resistance Range Selected	Resistance in $\mu\Omega$	Acceptance limits of Resistance in $\mu\Omega$		Test Results Resistance in $\mu\Omega$	
			MIN	MAX	CH1	CH2
1	1000 $\mu\Omega$	160	156.7	163.3	159.1	158.3
		1600	1567	1633	1592	1601
2	2000 $\mu\Omega$	160	156.7	163.3	160.5	161.1
		1600	1567	1633	1603	1609
3	4000 $\mu\Omega$	160	156.7	163.3	161.1	159.3
		1600	1567	1633	1613	1612
4	8000 $\mu\Omega$	160	156.7	163.3	159.6	160.2
		1600	1567	1633	1614	1602

2.0 BRIEF RESULTS:

The product under test complies with calibration specification indicated above and is verified to be within calibration.

Tested By	Certified By
Akash Sirsat	 