



# Precise Calibrations (P) Ltd.

(N.A.B.L. Accredited Laboratory)  
एन.ए.बी.एल. द्वारा प्रमाणित प्रयोगशाला  
Electro-Technical Calibration Services  
विद्युत तकनीकी अंशांकन सेवाएं

Basement, Plot No. 85, Sector - 5, Rajendra Nagar, Sahibabad, Ghaziabad (U.P.) - 201005  
Ph : +91 0120-4106400, +91 8744050381  
Email : precisecalibrations@gmail.com, calibrations@evertechgroup.com



## CALIBRATION CERTIFICATE

ULR No. :	CC301223000000082F	Calibration Date :	09-05-2023
Certificate No.:	PCPL/22-23/05/0003-09	Calibration Recommended Date:	08-05-2024
Location:	Lab	Certificate Issue Date:	10-05-2023
Service Request No.:	PCPL/SRF/22-23/05/0003	Service Request Date:	09-05-2023
Condition on Receipt:	OK	Date of Receipt:	09-05-2023

### CUSTOMER DETAILS

Name & Address: M/s. Oasis Techno Engineers Pvt. Ltd.  
B-42, 2nd Floor, Sec-65, Noida - 201 301, U.P.

### DESCRIPTION OF INSTRUMENT FOR CALIBRATION

Name of Instrument : DC High Voltage Test Set (0-70kV)  
Make: RE  
Model/Type: -----  
Serial No./ Instrument ID: 8890-198(P)-10-11


### Environmental Conditions

Temperature : 25±4 °C Humidity: 30% to 75%

### DETAILS OF MASTER INSTRUMENTS

Name of Reference Standard: HV Probe With DMM  
Make & Model: Fluke/80K-40, Rishabh/14S  
Serial No.: 29650008, 1070807612  
Traceable To: C & I Calibrations, Jaipur  
Calibrated Parameter: High Voltage  
Work Instruction No.: PCPL/WI-01/006  
Reference Standard: IS 13875

Note: All the instruments used are traceable to national standard through reference standards and their calibrations are valid.



Calibrated By:	<i>Yogesh</i>	Issued By:	<i>Barekhat</i>
	Calibration Engineer		Authorized Signatory



# Precise Calibrations (P) Ltd.

(N.A.B.L. Accredited Laboratory)  
एन.ए.बी.एल. द्वारा प्रमाणित प्रयोगशाला  
Electro-Technical Calibration Services  
विद्युत तकनीकी अंशांकन सेवाएं

Basement, Plot No. 85, Sector - 5, Rajendra Nagar, Sahibabad, Ghaziabad (U.P.) - 201005  
Ph : +91 0120-4106400, +91 8744050381

Email : precisecalibrations@gmail.com, calibrations@evertechgroup.com



ULR No. :	CC301223000000082F	Calibration Date :	09-05-2023
Certificate No.:	PCPL/22-23/05/0003-09	Calibration Recommended Date:	08-05-2024

## ELECTRO-TECHNICAL CALIBRATION RESULTS

### 1. DC High Voltage Measurement :-

Sr. No.	Set Value In *UUC (V)	Observed Value In Master (1V = 1kV)	(±) Expanded Uncertainty (%)
1	5	5.02	3.5
2	10	10.08	3.5
3	15	15.13	3.5
4	20	20.19	3.5
5	25	25.22	3.5
6	28	28.27	3.5

Note:- In Master Readings are 1V= 1 kV

### Remarks:

- 1) \* UUC: Unit under Calibration.
- 2) This Certificate refers only to the particular item submitted for the calibration.
- 3) The confidence level is 95% (approximately.) at k=2.
- 4) The results reported in this Certificate are Valid at the time and under the stipulated conditions of measurement in laboratory only.
- 5) This certificate shall not be reproduced, except in full, unless written permission for the publication of an approved abstract has been obtained from the Director of M/s Precise Calibrations Pvt. Ltd.
- 6) PCPL is responsible for all the information in the calibration certificate, except that provided by the customer, information supplied by the customer can affect the validity.

\*\*\* End of Certificate \*\*\*



Calibrated By:		Issued By:	
	Calibration Engineer		Authorized Signatory





# Precise Calibrations (P) Ltd.

(N.A.B.L. Accredited Laboratory)  
एन.ए.बी.एल. द्वारा प्रमाणित प्रयोगशाला  
Electro-Technical Calibration Services  
विद्युत तकनीकी अंशांकन सेवाएं

Basement, Plot No. 85, Sector - 5, Rajendra Nagar, Sahibabad, Ghaziabad (U.P.) - 201005  
Ph : +91 0120-4106400, +91 8744050381  
Email : precisecalibrations@gmail.com, calibrations@evertechgroup.com




## CALIBRATION CERTIFICATE

ULR No. :	CC30122300000083F	Calibration Date :	09-05-2023
Certificate No.:	PCPL/22-23/05/0003-10	Calibration Recommended Date:	08-05-2024
Location:	Lab	Certificate Issue Date:	10-05-2023
Service Request No.:	PCPL/SRF/22-23/05/0003	Service Request Date:	09-05-2023
Condition on Receipt:	OK	Date of Receipt:	09-05-2023

CUSTOMER DETAILS		
Name & Address:	M/s. Oasis Techno Engineers Pvt. Ltd. B42, 2nd floor , Sector -65 Noida, UP-201301	
DESCRIPTION OF INSTRUMENT FOR CALIBRATION		
Name of Instrument :	CT Analyzer	
Make:	Omicron	
Model	Digital	
Serial No./ Instrument ID:	NG668M	
Environmental Conditions		
Temperature :	25±4 °C	Humidity: 30% to 75%
DETAILS OF MASTER INSTRUMENTS		
Name of Reference Standard:	5.5 DMFC, 6.5 DMM	
Make & Model:	Zeal/ZSNMFC, Fluke/8846A	
Serial No.:	201202418, 2510012	
Traceable To:	PCP, AIPL	
Work Instruction No.:	PCPL/WI-01/001 ,PCPL/WI-01/008	
Reference Standard:	IS-13875	

Note: All the instruments used are traceable to national standard through reference standards and their calibrations are valid.



Calibrated By:	<i>yogesh</i>	Issued By:	<i>Barukul As</i>
	Calibration Engineer		Authorized Signatory